

Data Sheet: CBC Tourniquet



Material	
Casing	Polycarbonate (PC)
Further casing parts	Polyoxymethylene (POM)
Ribbon	Polyester and Lycra [®] , phthalate-free
Ribbon colors	The colors do NOT contain the following chemical substances which are prohibited in textiles: 4-Aminobiphenyl, Benzidine, 4-Chloro-o-toluidine, 2-Naphthylamine, o-Aminoazotoluene, p-Chloraniline, 2-Amino-4-nitrotoluene, 2,4-Diaminoanisole, 3,3'-Dichlorobenzidine, 3,3'-Dimethoxybenzidine, 3,3'-Dimethylbenzidine, 4,4-Diaminodiphenylmethane, p-Cresidine, 4,4-Methylenebis(2-chloroanilin), 4,4'-Oxydianiline, 4,4'-Thiodianiline, o-Toluidine, 2,4-Diaminotoluen, 2,4,5-Trimethylaniline.

Specifications to maintain hygiene	
Classification	The tourniquet can be classified as uncritical medical product, provided that it is applied on intact skin. Do not apply on pathologically altered skin.

Cleaning procedure	
Washing	With minor staining, the tourniquet can be washed by hand with domestic laundry detergents at up to 65 °C. It is recommended to use hand wash detergents which offer a good cleansing effect at a temperature of 40 °C. Please observe the respective manufacturer's instructions of use. Please rinse the tourniquet at least twice with fresh, lukewarm water. Avoid scented detergents. Due to the variety of detergents available in the market, we cannot recommend a specific product.
Note	If the tourniquet is contaminated with blood, secretion, or other body fluids, please dispose of it immediately in accordance with all pertinent regulations on the disposal of biohazardous waste. In this case, it is not permissible to clean and reuse the tourniquet.

Procedure of disinfection	
Physical procedure	Boiling in water / steam sterilization
Note	Physical disinfection procedures are not recommended by the manufacturer, but the facility may validate their own disinfection protocol.

Chemical procedure	For our tourniquet, we recommend a disinfection with a liquid disinfectant. Use CIDEX OPA by Medos International Sarl as disinfectant. Place the tourniquet in a container. Do not use a metallic container. Do not coil up the tourniquet. Pour as much disinfectant solution into the container to completely cover the tourniquet and leave it for at least 5 minutes in the solution. The room temperature shall be at least 20 °C. After the contact time, rinse the tourniquet with plenty of fresh water with at least 3 rinsing rounds each with fresh water. Leave the tourniquet in the rinse water for at least 1 minute. You can hang up the tourniquet for drying. Make sure that dripping water does not do any damage. For more information on safe use, disposal and storage life of the disinfectant, refer to the corresponding instructions of use. <u>Do not use</u> disinfectants containing chlorine-based release agents or aromatic hydrocarbons.	
Manufacture and distribution of CIDEX OPA	Manufacturing: Medos International Sarl Chemin-Blanc 38 2400 Le Locole, Switzerland	Distribution in Germany: Johnson+Johnson Medical GmbH Oststraße 1 22844 Norderstedt