

# Evaluation of VACUETTE<sup>®</sup> K<sub>3</sub>EDTA Evacuated Blood Collection Tubes for Immunoematology

**Background:**

Greiner Bio-One, Austria has sold plastic evacuated tubes (VACUETTE<sup>®</sup>) for venous blood collection since 1986.

The Greiner VACUETTE<sup>®</sup> K<sub>3</sub>EDTA tube provides a means of collecting and transporting an undiluted plasma specimen in a closed evacuated system. The tube contains spray-dried EDTA, yielding a ratio of 1.8mg/mL of blood when the evacuated tube is filled correctly to its fill volume. The EDTA binds calcium ions which block the coagulation cascade.<sup>1,2</sup>

VACUETTE<sup>®</sup> EDTA tubes are used for testing whole blood in the clinical laboratory and may be used for testing in routine immunoematology i.e. red cell grouping, Rh typing and antibody screens.

**Study Objective:**

A clinical evaluation was carried out to compare the performance of the Greiner Bio-One VACUETTE<sup>®</sup> K<sub>3</sub>EDTA tubes to the Becton Dickinson Vacutainer<sup>®</sup> Glass K<sub>3</sub>EDTA tube and Vacutainer<sup>®</sup> PLUS K<sub>2</sub>EDTA tube.

**Study Design:**

The study design was based on recommendations made by reviewers from the FDA Center for Biologics Evaluation and Research. Division of blood Applications (CBER).

*The following tube types were used in this study:*

Sample No.	Description
1	Greiner VACUETTE <sup>®</sup> K <sub>3</sub> EDTA, 6 mL, 13x100mm tube
2	Greiner VACUETTE <sup>®</sup> K <sub>3</sub> EDTA, 3 mL, 13x75mm tube
3	Becton Dickinson Vacutainer <sup>™</sup> Glass K <sub>3</sub> EDTA, 7 mL, 13x100mm tube, (comparator device)
4	Becton Dickinson Vacutainer <sup>™</sup> PLUS K <sub>2</sub> EDTA, 3 mL, 13x75mm tube, (comparator device)

Blood specimens were obtained using each site's standard phlebotomy techniques referencing Standard Operating Procedures and OSHA's safety requirements for blood collection. The order of draw was randomized.

The following two tubes were drawn from each donor at two Donor Centers:

- 1) one Greiner VACUETTE<sup>®</sup> K<sub>3</sub>EDTA, 6 mL, 13x100mm tube, and
- 2) one Becton Dickinson Vacutainer<sup>®</sup> Glass K<sub>3</sub>EDTA, 7 mL, 13x100mm tube.

In addition, one Greiner 13x100 mm VACUETTE<sup>®</sup> K<sub>3</sub>EDTA, 6 mL half evacuated (3 mL) tube was collected from each of the 10 known red cell antibody positive donors at Donor Center #1 to simulate partial draw.

The following two tubes were drawn from each patient at the University Hospital:

- 1) one Greiner VACUETTE<sup>®</sup> K<sub>3</sub>EDTA, 3 mL, 13x75mm tube and
- 2) one Becton Dickinson Vacutainer<sup>®</sup> PLUS K<sub>2</sub>EDTA, 3 mL, 13x75mm tube.

**A. Donor Center – Site #1:**

- 1) 50 apparently healthy donors (full draw tubes)
- 2) Subset: 10 apparently healthy donors for antigen phenotyping
- 3) Subset: 10 apparently healthy donors for delayed antigen phenotyping (0, 15 or 19 days)
- 4) 15 known red cell antibody positive blood donors (full draw tubes)
- 5) Subset: 10 known antibody positive donors (partial draw/ half-evacuated tubes)
- 6) Subset: 10 known antibody positive individuals (full and partial draw/ half-evacuated tubes) for delayed testing

**B. Donor Center- Site #2:**

The following donors were drawn:

- 1) 52 apparently healthy donors
- 2) Subset: 10 apparently healthy donors for antigen phenotyping
- 3) 10 known red cell antibody positive donors

**C. University Hospital – Site #3:**

Fifty patients, as follows:

- 1) Multi-transfused [Hb SS (2), thalassemia (1) and others with antibodies (2)] (5)
- 2) Cardiology (5)
- 3) Leukemia (5)
- 4) Bone Marrow Transplant (5)
- 5) Liver Disease (5)
- 6) General Surgery (10)
- 7) General Medicine (15)

The tubes were gently mixed using eight complete inversions immediately following blood collection. Tubes were centrifuged using the laboratory's standard procedure, to separate cellular elements completely from the plasma.

All but three samples were tested at each of the Donor Centers within 24 hours. At Donor Center #1, testing was delayed for two days for two positive antibody samples and three days for another positive antibody sample. In addition, ninety-five percent of the samples at the University Hospital were tested within eight hours of collection, while the remaining five percent were tested within 16 hours of collection.

The following tests and instrumentation and tests were used:

**A. Donor Center – Site #1:**

- Olympus® PK7200™ Automated Microplate System: ABO, Rh
- Manual Method: DAT, Antibody Screening and Identifications, Antigen Phenotyping
- Sample Stability Study/Delay in Testing: 1) Antibody Positive Samples: ABO, Rh, Antibody Screening and Identification using full and partial draw/ half-evacuated tubes, and 2) Antigen Phenotyping Samples: Antigen Phenotyping using full draw tubes.

**B. Donor Center – Site #2:**

- Ortho ID-Micro Typing System™ (ID-MTS) Gel Test™: ABO, Rh, DAT, Antibody
- Screening and Identifications, Antigen Phenotyping
- Ortho Selectogen® Reagent Red Blood Cells Two Cell Panel: Antibody Screening
- Ortho's Resolve® Panel A: Antibody Panels and Identifications

**C. University Hospital – Site #3:**

- Immucor® ABS2000: ABO, Rh, Antibody Screening
- Manual Method: ABO, Rh, DAT
- Standard LISS Tube Method: Antibody Screening and Identification

**Conclusion:**

The Greiner VACUETTE® K<sub>3</sub>EDTA tubes (full and partial draw/ half-evacuated) demonstrated substantial equivalence to the Becton Dickinson Vacutainer® Glass K<sub>3</sub>EDTA and PLUS K<sub>2</sub>EDTA tubes with various standard assays using donor and recipient populations. Antigen and antibody identification did not change over time when samples were stored in the Greiner VACUETTE® K<sub>3</sub>EDTA tubes, demonstrating that these proteins were not adsorbed onto the plastic walls of the tubes and interfering substances were not leached from the walls of the tubes.<sup>3,4,5,6,7,8,9</sup>

**Results/Discussion:**

**ABO/Rh Typing**

ABO/Rh typing was performed on matching tubes of blood from 102 apparently healthy donors, 25 known antibody positive donors and 50 patients. The testing was performed using an Olympus® PK7200™, an Immucor® ABS2000 and the ID-MTS Gel Test™, according to each manufacturer's recommended procedure.

The positive antibody donors at Donor Center #1 had the ABO and Rh typing performed manually. In addition, ABO/Rh typing was manually performed on the fifty patients, in parallel with the Immucor® ABS2000 testing. There were no inaccurately reported results with the Greiner VACUETTE® K<sub>3</sub>EDTA tubes when compared to the BD Vacutainer® EDTA tubes.

**Antigen Phenotyping**

Antigen Phenotyping was performed on matching tubes of blood from 20 apparently healthy blood donors. The samples were screened for the most common antigens of the Rh (C, E, c, e), Kell (K), Duffy (Fy<sup>a</sup>, Fy<sup>b</sup>), Kidd (Jk<sup>a</sup>, Jk<sup>b</sup>), and MNS (M, N, S, s) blood group systems. The distribution of results is summarized in Table #1.

Table #1		
	Donor Center – Site #1 (#Pos/#Neg)	Donor Center – Site #2 (#Pos/#Neg)
C	6/4	5/5
E	1/9	2/8
c	8/2	NT
e	9/1	NT
K	0/10	1/9
k	NT	10/0
Fy <sup>a</sup>	7/3	6/4
Fy <sup>b</sup>	9/1	5/5
Jk <sup>a</sup>	8/2	9/1
Jk <sup>b</sup>	5/5	7/3
S	7/3	4/6
s	10/0	10/0
M	7/3	NT
N	8/2	NT

\*NT = Not tested

**Antibody Screening and Identification**

**Full Draw Tube**

Antibody Screening was performed on 102 apparently healthy blood donors, 25 known positive blood donors, and 50 patients using the full draw Greiner VACUETTE® K<sub>3</sub>EDTA tube and the BD Vacutainer® EDTA tubes. The testing was performed using the Immucor® ABS2000, the ID-MTS Gel Test™, or a manual system according to the manufacturer's recommended procedures. Antibody screening was manually performed on the 50 patients in parallel with the Immucor® ABS2000 testing. All positive antibody screening samples were followed up with antibody identification.

Concordant results were obtained between the Greiner VACUETTE® K<sub>3</sub>EDTA tubes when compared to the BD Vacutainer® EDTA tubes.

However, in some of the comparisons, there was a 1+ difference in reaction grade, but none of these results demonstrated a change to a negative reading.

This variation is within the expected reproducibility of a subjective grading system.

#### Partial-Draw Tube

In addition, ABO/Rh typing, antibody screening and antibody identification were performed on a subset of 10 of the known antibody positive blood donors using partial draw/ half-evacuated Greiner VACUETTE® K<sub>3</sub>EDTA tubes and full draw BD Vacutainer® Glass K<sub>3</sub>EDTA tubes. The testing was performed manually, according to the Donor Center's established procedure.

Concordant results were obtained between the partial draw/ half-evacuated Greiner VACUETTE® K<sub>3</sub>EDTA tubes and the full draw BD Vacutainer® Glass K<sub>3</sub>EDTA tubes. However, in some of the comparisons, there was a 1+ difference in reaction grade, but none of these results demonstrated a change to a negative reading. This variation is within the expected reproducibility of a subjective grading system.

#### Delay in Testing

Ten of the antigen phenotyping samples and 10 of the known antibody positive blood donor samples (full and partial draw/ half-evacuated tubes) were stored at 2-8°C following initial testing. Testing was repeated at 15-19 days after collection. The antigen phenotyping samples were only repeated for antigen phenotyping testing. The antibody positive blood donor samples were repeated for ABO/Rh typing and antibody screening and identification.

Concordant results were obtained between the full and partial draw/ half-evacuated Greiner VACUETTE® K<sub>3</sub>EDTA tubes and the full draw BD Vacutainer® Glass K<sub>3</sub>EDTA tubes at Day 14. However, in some of the comparisons, there was a 1+ difference in reaction grade. This variation is within the expected reproducibility of a subjective grading system. A decrease in grading results was observed in some samples between Day 0 the last day of testing (Day 15 or 19).<sup>10</sup> This is also not unexpected, considering the age of the sample.

#### DAT

Antibody screening was performed on 102 apparently healthy blood donors, 10 of the known positive blood donors, and 50 patients using the Greiner VACUETTE® K<sub>3</sub>EDTA tubes and the BD Vacutainer® EDTA tubes. There were no DAT positive results among the 112 blood donors and only one DAT positive result among the hospitalized patient's samples. Concordant results were obtained with the Greiner VACUETTE® K<sub>3</sub>EDTA tubes and the BD Vacutainer® EDTA tubes.

In addition, a panel of five simulated DAT positive samples were prepared and tested at Donor Center – Site #1 using the Greiner VACUETTE® K<sub>3</sub>EDTA tubes and the BD Vacutainer® Glass K<sub>3</sub>EDTA tubes.

Preparation of the coated red cells followed the procedure for using red cells coated with Anti-Fy<sup>a</sup> described in the FDA Center for Biologics Evaluation and Research Guidance Document "Recommended Methods for Anti-Human Globulin Evaluation", issued in March 1992.<sup>11</sup> The dilutions used in this study were selected to represent a range of positive reactivity. The samples were tested on Day 0 (date of preparation) and repeated on Days 7 and 14. Concordant results were obtained between the Greiner VACUETTE® K<sub>3</sub>EDTA tubes and the BD Vacutainer® Glass K<sub>3</sub>EDTA tubes on Days 0, 7 and 14. In some of the samples, there was a 1+ difference in reaction grade of the results. This variation is within the expected reproducibility of a subjective grading system.

#### References:

1. Greiner Bio-One. Evacuated Blood Collection System For In Vitro Diagnostic Use. Product Insert. Kremsmünster, Austria. 2001.
2. Gruber, H. Greiner Bio-One. Product Manual. Kremsmünster, Austria. July 2002.
3. Greiner Bio-One 510(k) Submission. Blood Collection Tube – EDTA K3 Pre-Market Notification Addition of Immunohematology Claim. Monroe, NC. December 2002.
4. Kemper, M. Final Report: Greiner VACUETTE® Evacuated Blood Collection Tubes Blood Bank Study K<sub>2</sub>EDTA, K<sub>3</sub>EDTA vs. BD K<sub>3</sub>EDTA. SMF-Center for Blood Research. Sacramento, California. March 14, 2003.
5. Kemper, M. Personal Communication. SMF-Center for Blood Research. Sacramento, California. March 2003.
6. Korte, L. Final Report: Comparison of Greiner VACUETTE Blood Collection Tubes. Gulf Coast Regional Blood Center. Houston, Texas. January 23, 2003.
7. Korte, L. Personal Communication. Gulf Coast Regional Blood Center. Houston, Texas. March 2003.
8. Langeberg, A. Personal Communication. Georgetown University Hospital. March 24, 2003.
9. Sandler, G.S., M.D. Personal Communication. Georgetown University Hospital. March 2003.
10. Sandler, G.S., M.D. Personal Communication. Georgetown University Hospital. July 2003.
11. FDA Center for Biologics Evaluation and Research Guidance Document. Recommended Methods for Anti-Human Globulin Evaluation. March 1992.

VACUETTE is a registered trademark of Greiner Bio-One. Vacutainer is a trademark of Becton, Dickinson and Company. ID-Micro Typing System and Gel Test are trademarks of Ortho-Clinical Diagnostics. Selectogen and Resolve are registered trademarks of Ortho-Clinical Diagnostics. Immucor is a registered trademark of Immucor, Inc. Olympus is a registered trademark and PK7200 is a trademark of Olympus America, Inc.