making a difference

INSTRUCTIONS FOR USE

CELLDISC Multi Layer Device (1 - 40 Layers)





www.gbo.com

SIMPLIFY YOUR CELL CULTURE WORK PROCESS

CELLDISC MULTI LAYER DEVICE (1 - 40 LAYERS)

The Greiner Bio-One CELLdisc is a ready-tostart, multi layer device, as easy to use as a T-flask. The innovative ergonomic CELLdisc design provides a versatile system for the propagation of adherent mammalian cells from research scale to industrial batches. It is available either with the standard tissue culture surface (TC; red screw cap) or the Advanced TC surface (Adv. TC, blue screw cap) identical to all Greiner Bio-One cell culture products to assure consistent performance from lot to lot and from format to format.

INTENDED USE

General laboratory products for cell culture to be used by qualified personnel in a laboratory environment.



Figure 1: CELLdisc with 1, 16 and 40 layers

CELLDISC SINGLE LAYER

The single layer CELLdisc (CD1), in contrast to the larger multi layered versions, allows easy monitoring of cell growth, cell morphology and confluency. As conditions within the single layer CELL disc will be identical to those within larger units, the CD1 can be used to anticipate when media changes are required and harvesting is recommended in any larger units being cultivated alongside the CD1. To ensure that the ambient conditions for the reference CD1and the multilayer CELL disc are absolutely identical, the CELLevator allows storage of the single layer CELLdisc on top of the CELLdisc production unit. To maintain the accustomed quality and efficiency of cell culture, customers are advised to test Greiner Bio-One systems under the conditions defined in their own protocols.



Figure 2: General CELLdisc details

1/ SINGLE LAYER CELLDISC

Unpack the single layer CELLdisc and place it in a laminar air flow cabinet in order to work in sterile conditions.

Prepare cell suspension in accordance with the concentration (cells/cm²) used with other disposables for adherent cell culture.

Unscrew screw cap and transfer the cell suspension directly into the CELLdisc using the large opening port either by pouring or pipetting.

Firmly tighten the screw cap onto the CELLdisc to close it.

Tilt the single layer CELLdisc gently from one side to the other to assure that media and cells distribute evenly.



WARNING

The media or cell suspension should not touch the filter. If the filter has absorbed any fluid, this will inhibit any gas transfer into and out of the CELLdisc. In this case the disposable has to be discarded and a new single layer CELLdisc has to be used.

2/ CELLDISC 4 - 24 LAYERS



Unpack the CELLdisc and place it in a laminar air flow cabinet in order to work in sterile conditions.

Prepare cell suspension in accordance with the concentration (cells/cm²) used with other disposables for adherent cell culture. Hold the CELLdisc at an angle of 30°.

Hold the CELLdisc with the screw cap at a position of approximately 105° for right-handed users or 255° for left-handed ones.

To simplify filling procedure, the CELLdisc can be positioned on the CELLstage filling device (see ordering information).

$\underline{\land}$

WARNING

To obtain equivalent cell growth in all layers, formation of air bubbles must be avoided during CELLdisc processing. Therefore, an exact angle of 30° and a specific position of the connecting channel must be maintained during filling of 4-24 layer CELLdisc. This handling procedure guarantees that the pressure is equalized through the connecting channel (indicated in blue in chapter 2, step 3) without contact with the filled in liquid. Thus, the air does not flow through the liquid and does not cause foaming. In addition to the filling process, any generation of air bubbles should be avoided. Vigorous shaking of the CELLdisc is not recommended. Larger volumes of liquids should be mixed outside the CELLdisc and then added to the disposable as described below. Small amounts can be pipetted directly into the CELLdisc and then distributed to all layers by repeating the equilibration process.



Unscrew screw cap and add the cell suspension either by pouring or pipetting directly into the CELLdisc using the large opening port.

The media will fill the topmost layer first and then move slowly to each layer underneath.

Wait until the liquid is distributed into the individual layers still keeping the CELLdisc in the indicated position before firmly tighten the screw cap to close it.

To start liquid equilibration lay down CELLdisc horizontally and turn it to assure that the media and all layers are in contact through the connecting channel.

$\underline{(}$

WARNING

The media or cell suspension should not touch the filter. If the filter has absorbed any fluid, this will inhibit any gas transfer into and out of the CELLdisc. In this case the disposable has to be discarded and a new CELLdisc has to be used.

2/ CELLDISC 4 - 24 LAYERS

NOTE: Small volumes of liquids RECOMMENDED HANDLING (e.g. trypsin) may accumulate in the upper layers during filling and not flow through the whole filling channel. To guarantee equal distribution, the liquid must be in contact with all layers through the filling channel before equilibration process is initiated. Therefore the CELLdisc must be positioned horizontally with the opening port at the lowest position. The media will now distribute equally between 6 all layers. Turn the CELLdisc to disconnect media flow from the filling channel. Do not rotate the CELLdisc any further as this could lead to contact between liquid and filter.

NOTE: During transport, tilt the CELLdisc slightly backward to assure that there is no liquid contact with the filling channel or accidental media flow to another layer.

To stack individual CELLdisc or a single layer CELLdisc on top of a production CELLdisc use the CELLevator.

From this position raise the CELLdisc upright

and place the disposable on a horizontal

Proceed with the cultivation based on the

surface inside an incubator.

appropriate protocol.

More details on CELLevator can be found on the reverse side.

For liquid removal, unscrew screw cap and tilt the CELLdisc slowly 90° with the large opening port at the lowest possible position and pour out the media or use a pipetting system.

2/ CELLDISC 4 - 24 LAYERS









•

3/ CELLDISC 40 LAYERS



As with the CELLdisc 4-24 layers, air bubble formation must be avoided during processing to obtain even cell growth in all layers. While the position of the screw cap/filling channel and the general filling procedure is identical to recommended handling of CD4-24 a smaller angle of 20° must be maintained during filling of a 40-layer CELLdisc.

This handling procedure guarantees that the pressure is equalized through the connecting channel (indicated in blue) without contact with the filled in liquid. Thus, the air does not flow through the liquid and does not create air bubbles.

To simplify filling procedure, the CELLdisc can be positioned on the CELLstage filling device (see ordering information).

To start liquid equilibration lay down the CELLdisc horizontally and turn it to assure that the media and all layers are in contact through the central filling channel.

The media will now separate equally between all layers.

the liquid must be in contact with all layers through the connecting channel before equilibration process is initiated. Therefore the CELLdisc must be positioned horizontally with the opening port

Turn the CELLdisc to disconnect media flow from the filling channel.

NOTE: To guarantee equal distribution,

at the lowest position.

From this position raise CELLdisc upright and place the disposable on a horizontal surface inside an incubator.

Liquid removal is identical to the recommended handling of CELLdisc 4-24 layers.

3/ CELLDISC 40 LAYERS



4/ CELLDISC 4 - 40 LAYERS

HARVESTING OF CELLS

The protocol for harvesting cells from a multilayer device such as CELLdisc which does not allow for direct pipette access varies slightly from your standard protocol. Especially as the detachment of cells can be visualized microscopically only for the bottom layer of a CELLdisc with one, four or maximum eight layers.

This cell harvest protocol refers to standard techniques and specific suggestions to gain maximum cell yields.

In general, we recommend using the same dissociating solution and concentrations for enzymatic detachment that is used to harvest these cells from a standard Greiner Bio-One cell culture disposable. As the surface treatments and basic materials are identical for these vessels, standard protocols can be converted.

The only adaption required is based on growth area and cell numbers per layer. The use of chelating agents such as EDTA in addition to the enzyme (trypsin, papain etc.) may improve cellular detachment. Remove cultivation media either by pouring or aspiration as described above. Wash cells once with PBS or an equivalent buffer using approx. 20 ml per layer. Follow the same protocol as for CELLdisc filling to distribute the buffer through all layers. Then tilt the CELLdisc slowly back and forth to gently rinse each cell layer and remove all traces of media. Remove PBS either by pouring or aspiration. Thereafter add 7-10 ml of your enzymatic dissociation agent per layer. For a 4-layer CELLdisc this results in 28-40 ml which has to be equilibrated as described above.

Incubate CELLdisc at 37 °C and 5 % CO² for 3-5 minutes. Tapping CELLdisc from the side can accelerate cellular detachment. Strong adherent cells might need longer incubation or stronger tapping. Stop the enzymatic reaction by adding the same amount of serum-containing media or an appropriate inhibitor to the CELLdisc, equilibrate the liquids and mix the solutions by gently tilting the CELLdisc back and forth. Thereafter cell suspension can be harvested for further processing by pouring or aspiration.

If removal of the enzymatic dissociation agent is desired, then spin cell suspension at 100xg for 5 minutes. Remove the dissociation agent containing media and resuspend cell pellet in fresh media.



FOR FURTHER INFORMATION AND/OR SAMPLE ORDERING PLESAE VISIT OUR WEBSITE OR CONTACT US.

CELLdisc with TC surface

Sterility SAL 10⁻⁶

ltem No.	Layers	Surface	Growth area [cm³]	Min. working volume[ml]	Max. working volume[ml]	Ø[mm]	Bag[pcs]
678101	1	• TC	250	15	70	200	8
678104	4	• TC	1.000	60	280	200	4
678108	8	• TC	2.000	120	560	200	3
678112	12	• TC	3.000	180	840	200	2
678116	16	• TC	4.000	240	1.120	200	2
678124	24	• TC	6.000	360	1.680	200	2
678140	40	• TC	10.000	600	2.800	200	1

CELLdisc with TC surface

Sterility SAL 10⁻⁶

ltem No.	Layers	Surface	Growth area [cm³]	Min. working volume[ml]	Max. working volume[ml]	Ø[mm]	Bag[pcs]
678901	1	🔵 Adv. TC	250	15	70	200	8
678904	4	🔵 Adv. TC	1.000	60	280	200	4
678908	8	🔵 Adv. TC	2.000	120	560	200	3
678912	12	🔵 Adv. TC	3.000	180	840	200	2
678916	16	🔵 Adv. TC	4.000	240	1.120	200	2
678924	24	🔵 Adv. TC	6.000	360	1.680	200	2
678940	40	🔵 Adv. TC	10.000	600	2.800	200	1

CELLdisc with external filter

Sterility SAL 10⁻⁶, triple packed

Sterility SAL 10 ⁻⁶ , triple packed			 Adv. TC on demanc 		
ltem No.	Layers	Surface	Description		
6781xx-EXF	1-40	• TC	CELLdisc with external filter		
	1-40	🔵 Adv. TC	CELL disc with external filter		

CELLdisc with closed filling caps

Sterility SAL 10 ⁻⁰ , triple packed			🔵 Adv. TC on deman		
Item No.	Layers	Surface	Description		
6781xx-CF1	1-40	• TC	CELLdisc with closed filling cap, single tubing and external filter		
6781xx-CF2	4-40	• TC	CELL disc with closed filling cap, double tubing and external filter		
	1-40	🔵 Adv. TC	CELLdisc with closed filling cap, single tubing and external filter		
	4-40	🔵 Adv. TC	CELLdisc with closed filling cap, double tubing and external filter		

CELLdisc Accessories

ltem No.	Description	Material	Measure [mm]	Weight[kg]
878072	CELLstage for CELLdisc with 4-24 layers	stainless steel	219x211x126	1.53
878073	CELLstage for CELLdisc with 40 layers	stainless steel	311x211x141	2.29

making a difference

www.gbo.com

GREINER BIO-ONE GMBH FRICKENHAUSEN, GERMANY

 PHONE
 +49 7022 948-0

 FAX
 +49 7022 948-514

 E-MAIL
 info@de.gbo.com

GREINER BIO-ONE IS A GLOBAL PLAYER. FIND THE CONTACT DETAILS OF YOUR



Devices of Greiner Bio-One are to be used by properly qualified persons only in accordance with the relevant Instructions for Use (IFU), where applicable. For more information contact your local Greiner Bio-One sales representative or visit our website (www.gbo.com).

All information is provided without guarantee despite careful processing. Any liability, warranty or guarantee of Greiner Bio-One GmbH is excluded. All rights, errors and changes are reserved. If not stated otherwise, Greiner Bio-One GmbH has all copyrights and/or other (user-)rights in this documents, in particular to signs such as the mentioned (word-picture-)brands and logos. Any use, duplication or any other use of the rights of Greiner Bio-One GmbH is expressly prohibited. **Media owner:** Greiner Bio-One GmbH / Represented by Managing Directors Jakob Breuer and Heinz Schmid. The company is registered in the Commercial Register at the first instance court in Stuttgart, HRB 224604 / VAT Number: DE812585719.

F000360 english [Rev.04, 09.2021]

