

**SAFE AND
CONSISTENT**



CLiP® PORTED

Automatic safety IV catheter

CLiP® Ported is an automatic safety IV catheter with injection port. CLiP® Ported is available in a comprehensive range to meet demanding physical conditions and different clinical needs. The products have been developed with the user and patient in mind - configured to provide excellent insertion properties. When patients and their conditions vary, CLiP® Ported remains consistent and its use contributes to safe and successful insertion.

- / Lightweight to not affect handling
- / Injection port for easy bolus administration
- / Blood sampling as indication for use
- / Suitable for high pressure injection up to 305 psi
- / Small environmental footprint with compact product and packaging design

ONLINE TRAINING FOR HEALTH CARE PROFESSIONALS.

We offer online courses to support health care professionals with best practice when working with CLiP® Ported automatic safety IV catheter.

Register for your course at stick-to-safety.com/training



PRODUCT OVERVIEW

FOR FURTHER INFORMATION AND/OR SAMPLE ORDERING
PLEASE VISIT OUR WEBSITE OR CONTACT US.

CLiP® Ported Automatic safety IV catheter

Not made with natural rubber latex and DEHP and PVC free

Item no. PUR	Item no. FEP	Catheter size	Catheter dimension	Colour code*	Catheter flowrate	Qty. inner / outer
VP144511	VP144501	14 G	2.0 x 45	● orange	290	50 / 1,000
VP164511	VP144501	16 G	1.7 x 45	● grey	200	50 / 1,000
VP174511	VP174501	17 G	1.5 x 45	○ white	140	50 / 1,000
VP184511	VP184501	18 G	1.2 x 45	● green	100	50 / 1,000
VP183211	VP183201	18 G	1.2 x 32	● green	110	50 / 1,000
VP203211	VP203201	20 G	1.0 x 32	● pink	64	50 / 1,000
VP202511	VP202501	20 G	1.0 x 25	● pink	68	50 / 1,000
VP222511	VP222501	22 G	0.8 x 25	● blue	38	50 / 1,000
VP241911 ^[1]	VP241901 ^[1]	24 G	0.7 x 19	● yellow	22	50 / 1,000

PUR = Polyurethane
FEP = Fluorinated Ethylene Propylene

* Colour-coded according to ISO10555-5:2013

RECOMMENDED HANDLING

Please observe the instructions for use at www.stick-to-safety.com or in the shelf box.



1

Insert the catheter and observe the blood response in the flashback chamber, or between the needle and catheter for 24 Gauge ^[1].



2

Occlude the vein and withdraw the needle straight back. Observe safety clip activation.



3

Fixate and flush the catheter according to local protocol. Cover the site with a sterile dressing.

^[1]With needle notch for instant blood response between needle and catheter in front of the wing housing, not in the chamber behind the wing housing