

Certificate of Quality

REF

XXXXXX

LOT

XXXXXXXX

Product description: TEXT
TEXT
TEXT
TEXT

Following inspections were carried out and passed:

raw material	<input checked="" type="checkbox"/>	passed
visual	<input checked="" type="checkbox"/>	passed
physical	<input checked="" type="checkbox"/>	passed

Free of detectable DNase/RNase/human DNA
Representative product was tested by Real-Time-PCR.

Detection limits:

DNase	1.3×10^{-8} Kunitz units
RNase	5.0×10^{-11} Kunitz units
hDNA	5 pg

Free of detectable endotoxins (pyrogens)
Representative product was tested according to the United States Pharmacopoeia (USP) Chapter <85>
Detection limit: 0.03 EU/ml

Non-cytotoxic
Representative product was tested according to ISO 10993-5 and has been shown to be non-cytotoxic.

We hereby certify that our products were manufactured and tested in accordance with the following international standards and meet the required specifications:

ISO 10993-5 "Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity"

ISO 11137 "Sterilization of health care products - Radiation"

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.



yyyy-mm-dd

STERILE

SAL 10^{-6}

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.



Certificate creation date: yyyy-mm-dd

www.gbo.com