

CONFIRMATION

Transportation by aircraft

Product description: Cryo.s™, Cryo.s™ with Datamatrix

Item no.: 12x2xx, 12x2xx-xxx,

This Article meets the *pressure* requirements for *transportation by aircraft*

(Pressure requirements: packaging must be designed and constructed to prevent leakage that may be caused by changes in altitude and temperature during transportation aboard aircraft)

The hydrostatic pressure test was performed based on the ICAO.IATA DGR, 42nd Edition, Sections 6.3.5 and 5.0.2.9 and Title 49 CFR, Section 178.605

A sample of cryo vials was hydrostatically tested to 95 Kpa pressure differential as required for air shipment.

All tested cryo vials passed the test without any leakage

Please note that this confirmation expires after 12 months or in case of regulatory changes.



This document has been created digitally and is valid without a signature.
Confirmation creation date: 2018-07-03