

VACUETTE® Luer Adapter Instructions for use

Intended Use

The **VACUETTE®** Luer Adapter is for use in venous blood collection with attached butterfly needle or luer needle.

Product Description

The **VACUETTE®** Luer Adapter is a sterile, single-use device, manufactured from stainless steel and polypropylene and is fitted with a safety valve at one end. The perforated label not only serves to simplify identification, but also acts as a seal of integrity.

The product is not made with natural rubber latex and is non-DEHP.

Procedure

NOTE: If the paper label is broken, discard and use another luer adapter.

1. Grasp both ends of the luer case and carefully twist. The perforation of the seal will now break as the cap is twisted.
2. Remove the short end of the cover.
3. Thread the needle end of the luer adapter into a holder. Make sure the luer adapter is securely seated.
NOTE: It is recommended to use the **VACUETTE®** Luer Adapter with **VACUETTE®** Holders.
4. Remove the shield from the luer adapter and attach a winged needle or luer needle.
5. Perform venipuncture and allow all tubes to fill sufficiently.
NOTE: Always hold the tubes in place to ensure complete vacuum draw.
6. After venipuncture, the holder + luer adapter + puncture device should be disposed of together as a unit in a sharps container according to your facility's protocol.

Precautions/Cautions

- The handling of **VACUETTE®** Luer Adapter must be in accordance to the policies and procedures of your facility.
- HBV, HCV, HIV and other infectious diseases can be transmitted through contact with biological samples. In the event of exposure to biological samples, ensure that appropriate medical attention is obtained.
- Discard all **VACUETTE®** Luer Adapters in biohazard containers approved for their disposal.
- Do not use **VACUETTE®** Luer Adapters after their expiry date.
- Wear gloves during venipuncture and when handling blood collection tubes to minimise exposure hazard.
- The product has not been validated for use with arterial or venous catheters.
- A re-use of the device can lead to harmful injuries.
- Special care has to be taken when used on pregnant or nursing women and children.
- This device is to be used by trained healthcare professionals only in accordance with these instructions.
- The fluid pathway of the chosen accessory must be free of clots as this could block the **VACUETTE®** Luer Adapter.

Storage of VACUETTE® Luer Adapter before use

Recommended storage temperature: 4–36°C (40–97°F).

NOTE: Avoid exposure to direct sunlight. Exceeding the maximum recommended storage temperature may lead to impairment of the **VACUETTE®** Luer Adapter quality.

Label information

	Manufacturer		Temperature limit
	Use-by date		Do not re-use
	Batch code		Sterilized using ethylene oxide
	Catalogue number		Consult instructions for use
Rx only	Prescription device		

Production location:
Nipro Medical Industries Ltd.
Tatebayashi Plant, 2-19-64, Matsubara,
Tatebayashi-shi, Gunma, 374-8518, Japan
Made in Japan
Distributed by Greiner Bio-One GmbH, Austria



Greiner Bio-One GmbH
Bad Haller Str. 32
4550 Kremsmünster
Austria

www.gbo.com/preanalytics
office@at.gbo.com
Phone +43 7583 6791